

HEMOCHRON SIGNATURE ELITE DAILY MAINTENANCE AND QUALITY CONTROL RECORD

YEAR: _____

AREA: _____

MONTH: _____



				Temp Probe Result	EQC Level 1	EQC Level 2	Instrument ID:				
Day	Shift	Actual Time	10% Bleach	37±1.0 DEG	30±1 SEC	500±1 SEC	Normal Result	Abnormal Result	Operator ID	Comments	Lab Review
1	11PM-7AM										
1	7AM-3PM										
1	3PM-11PM										
2	11PM-7AM										
2	7AM-3PM										
2	3PM-11PM										
3	11PM-7AM										
3	7AM-3PM										
3	3PM-11PM										
4	11PM-7AM										
4	7AM-3PM										
4	3PM-11PM										
5	11PM-7AM										
5	7AM-3PM										
5	3PM-11PM										
6	11PM-7AM										
6	7AM-3PM										
6	3PM-11PM										
7	11PM-7AM										
7	7AM-3PM										
7	3PM-11PM										

Key: NIU = NOT IN USE, *Clean/Temp Probe/Electronic QC are done at start of each shift

HEMOCHRON SIGNATURE ELITE DAILY MAINTENANCE AND QUALITY CONTROL RECORD

YEAR: _____

AREA: _____

MONTH: _____



				Temp Probe Result	EQC Level 1	EQC Level 2	Instrument ID:				
Day	Shift	Actual Time	10% Bleach	37±1.0 DEG	30±1 SEC	500±1 SEC	Normal Result	Abnormal Result	Operator ID	Comments	Lab Review
8	11PM-7AM										
8	7AM-3PM										
8	3PM-11PM										
9	11PM-7AM										
9	7AM-3PM										
9	3PM-11PM										
10	11PM-7AM										
10	7AM-3PM										
10	3PM-11PM										
11	11PM-7AM										
11	7AM-3PM										
11	3PM-11PM										
12	11PM-7AM										
12	7AM-3PM										
12	3PM-11PM										
13	11PM-7AM										
13	7AM-3PM										
13	3PM-11PM										
14	11PM-7AM										
14	7AM-3PM										
14	3PM-11PM										

Key: NIU = NOT IN USE, *Clean/Temp Probe/Electronic QC are done at start of each shift

HEMOCHRON SIGNATURE ELITE DAILY MAINTENANCE AND QUALITY CONTROL RECORD

YEAR: _____

AREA: _____

MONTH: _____



				Temp Probe Result	EQC Level 1	EQC Level 2	Instrument ID:				
Day	Shift	Actual Time	10% Bleach	37±1.0 DEG	30±1SEC	500±1SEC	Normal Result	Abnormal Result	Operator ID	Comments	Lab Review
15	11PM-7AM										
15	7AM-3PM										
15	3PM-11PM										
16	11PM-7AM										
16	7AM-3PM										
16	3PM-11PM										
17	11PM-7AM										
17	7AM-3PM										
17	3PM-11PM										
18	11PM-7AM										
18	7AM-3PM										
18	3PM-11PM										
19	11PM-7AM										
19	7AM-3PM										
19	3PM-11PM										
20	11PM-7AM										
20	7AM-3PM										
20	3PM-11PM										
21	11PM-7AM										
21	7AM-3PM										
21	3PM-11PM										

Key: NIU = NOT IN USE, *Clean/Temp Probe/Electronic QC are done at start of each shift

HEMOCHRON SIGNATURE ELITE DAILY MAINTENANCE AND QUALITY CONTROL RECORD

YEAR: _____

AREA: _____

MONTH: _____



				Temp Probe Result	EQC Level 1	EQC Level 2	Instrument ID:				
Day	Shift	Actual Time	10% Bleach	37±1.0 DEG	30±1SEC	500±1 SEC	Normal Result	Abnormal Result	Operator ID	Comments	Lab Review
22	11PM-7AM										
22	7AM-3PM										
22	3PM-11PM										
23	11PM-7AM										
23	7AM-3PM										
23	3PM-11PM										
24	11PM-7AM										
24	7AM-3PM										
24	3PM-11PM										
25	11PM-7AM										
25	7AM-3PM										
25	3PM-11PM										
26	11PM-7AM										
26	7AM-3PM										
26	3PM-11PM										
27	11PM-7AM										
27	7AM-3PM										
27	3PM-11PM										
28	11PM-7AM										
28	7AM-3PM										
28	3PM-11PM										

Key: NIU = NOT IN USE, *Clean/Temp Probe/Electronic QC are done at start of each shift

HEMOCHRON SIGNATURE ELITE DAILY MAINTENANCE AND QUALITY CONTROL RECORD

YEAR: _____

AREA: _____

MONTH: _____



Instrument ID: _____

Day	Shift	Actual Time	10% Bleach	Temp Probe Result	EQC Level 1	EQC Level 2	Normal Result	Abnormal Result	Operator ID	Comments	Lab Review
29	11PM-7AM			37±1.0 DEG	30±1 SEC	500±1 SEC					
29	7AM-3PM										
29	3PM-11PM										
30	11PM-7AM										
30	7AM-3PM										
30	3PM-11PM										
31	11PM-7AM										
31	7AM-3PM										
31	3PM-11PM										
Key: NIU = NOT IN USE, *Clean/Temp Probe/Electronic QC are done at start of each shift											

Revision: 6-07