



Operating Instructions

863 Valley View Road, Eighty Four, PA 15330
Tel: 724-941-9701 Fax: 724-941-1369
www.skcinc.com

Preloaded Coated Filters Cat. No. 225-9014

SKC coated filters are shipped preloaded in the cassettes with end plugs and shrink bands in place. No assembly is required; they are ready to use.

Method & Chemical: OSHA ID-214 for Ozone

Filter & Coating:[†] 2 glass fiber filters; nitrite-impregnated. Three-piece 37-mm cassette. Closed-face sampling configuration

Prior to Sampling: Store at ≤ 39.2 F (4 C) for up to 6 weeks.

Sample Stability: Store at 71.6 F (22 C) for up to 30 days.

[†] Custom order due to very limited shelf-life.

**For information on other available
coated filters, request SKC Publication 1219.**

How to Use the SKC Preloaded Coated Filters

1. Select one coated filter cassette for calibrating the flow rate. A red plug secures the inlet, which is clearly marked "inlet," and a blue plug secures the outlet. Set up the sampling train for calibration with the representative filter cassette in line. For "closed-face" sampling, remove the plugs and connect the cassette to the sampling train. For "open-face" sampling, remove the outlet plug and the cassette inlet section then connect the cassette to the sampling train. For details on setting up a sampling train, refer to the SKC Sample Setup Guide "Sampling Train - Filters and Cyclones" (SKC Publication 1166).
2. Calibrate the pump to the desired flow rate using a film flowmeter or calibrated precision rotameter. For details on calibrating a pump, refer to the SKC Sample Setup Guide "Calibrating a Pump Using a Film Flowmeter" (SKC Publication 1163).
3. Replace the cassette used for calibration with a fresh coated filter cassette for sample collection. Attach the cassette to a worker's collar, as close to the breathing zone as possible, by using a filter cassette holder (SKC Cat. No. 225-1). Sample for the specified time interval and record the time.
4. Remove the filter cassette at the end of the specified sampling period and replace both end plugs (and the inlet if necessary). Recheck the flow rate using the same cassette and flowmeter used for calibration (in Step 2) to ensure that the flow rate has not changed by more than 5%.
5. Store samples at 71.6 F (22 C) for up to 30 days. Appropriately package samples, blanks, and all pertinent data to send to a lab for analysis. Ship samples via next day air to arrive at the lab on the following business day. Friday shipments are not recommended.

SKC Limited Warranty and Return Policy

SKC products are subject to the SKC Limited Warranty and Return Policy, which provides SKC's sole liability and the buyer's exclusive remedy. To view the complete SKC Limited Warranty and Return Policy, go to <http://www.skcinc.com/warranty.asp>.

Form 40043 Rev 1303