

Texas Department of Public Safety

BREATH TEST OPERATOR CERTIFICATION RECORD

Certificate Number **13949**

Operator Name **JAMES G DAVIDSON**

Technical Supervisor Area **037**

Agency **DPS PALMVIEW**

Address **722 BREYFOGLE RD.**
 PALMVIEW, TEXAS 78572



Date **10/23/2009**

Certification Status **CERTIFICATION RENEWED**


Technical Supervisor


Scientific Director

Texas Department of Public Safety

BREATH TEST OPERATOR CERTIFICATION RECORD

Certificate Number **13949**

Operator Name **JAMES G DAVIDSON**

Technical Supervisor Area **037**

Agency **DPS PALMVIEW**

Address **2401 N MOOREFIELD R**
 PALMVIEW, TEXAS 78572



Date **9/30/2008**

Certification Status **CERTIFICATION RENEWED**


Technical Supervisor


Scientific Director

Breath Test Operator Certification Record

Certificate Number **13949**
Operator Name **DAVIDSON, JAMES G**
Agency **DPS PALMVIEW**
Address **2401 N MOOREFIELD R**
PALMVIEW, TEXAS 78572
Service Code **THP**
Tech Supv Area **037**



PRACTICAL EXAMINATIONS

Instrument **INTOXILYZER 5000**

Practical Test 1

1. PASS
2. PASS
3. PASS
4. PASS
5. PASS

Practical Test 2

1. NP
2. NP
3. NP
4. NP
5. NP

Practical Exam **PASS**


WRITTEN EXAMINATION

Written Exam **PASS**

Effective Date **10/15/2007**

Certification Status **CERTIFICATION RENEWED**

Technical Supervisor


Marianela Martinez


OFFICE OF THE SCIENTIFIC DIRECTOR

Renewal Date **10/15/2007** Expiration Date **10/31/2008**

Comments:

Breath Test Operator Certification Record

Certificate Number **13949**
Operator Name **DAVIDSON, JAMES G**
Agency **DPS PALMVIEW**
Address **2401 N MOOREFIELD R**
PALMVIEW, TEXAS 78572
Service Code **THP**
Tech Supv Area **037**



PRACTICAL EXAMINATIONS

Instrument **INTOXILYZER 5000**

Practical Test 1	Practical Test 2
1. PASS	1. NP
2. PASS	2. NP
3. PASS	3. NP
4. PASS	4. NP
5. PASS	5. NP

Practical Exam **PASS**

WRITTEN EXAMINATION

Written Exam **N/R**

Effective Date **10/2/2006**

Certification Status **CERTIFICATION RENEWED**

Technical Supervisor

Mary Ann Peralez
Mary Ann Peralez

OFFICE OF THE SCIENTIFIC DIRECTOR

Renewal Date **10/2/2006** Expiration Date **10/31/2007**

Comments:

Breath Test Operator Certification Record

Certificate Number **13949**
Operator Name **DAVIDSON, JAMES G**
Agency **DPS PALMVIEW**
Address **2401 N. MOOREFIELD R**
PALMVIEW, TEXAS 78572
Service Code **THP**
Tech Supv Area **037**

PRACTICAL EXAMINATIONS

Instrument **INTOXILYZER 5000**

Practical Test 1	Practical Test 2
1. PASS	1. NP
2. PASS	2. NP
3. PASS	3. NP
4. PASS	4. NP
5. PASS	5. NP

Practical Exam **PASS**

WRITTEN EXAMINATION

Written Exam **PASS**

Effective Date **10/24/2005**

Certification Status **CERTIFICATION RENEWED**

Technical Supervisor 
Mary Ann Peralez

OFFICE OF THE SCIENTIFIC DIRECTOR

Renewal Date **10/24/2005** Expiration Date **10/31/2006**

Comments:

BREATH TEST OPERATOR CERTIFICATION RECORD

Certificate Number 13949
Operator Name DAVIDSON, JAMES G
Agency DPS PALMVIEW
Address 2401 N. MOOREFIELD R
PALMVIEW, TEXAS 78572
Service Code THP
Tech Supv Area 037

PRACTICAL EXAMINATIONS

Instrument INTOXILYZER 5000

Practical Test 1

1. PASS
2. PASS
3. NOT TAKEN
4. PASS
5. PASS

Practical Test 2

1. NOT TAKEN
2. NOT TAKEN
3. NOT TAKEN
4. NOT TAKEN
5. NOT TAKEN

Practical Exam PASS

WRITTEN EXAMINATION

Written Exam N/R

Effective Date 10/01/2004
Certification Status CERTIFICATION RENEWED
Technical Supervisor MaryAnn Peralez
MaryAnn Peralez

OFFICE OF THE SCIENTIFIC DIRECTOR

Renewal Date 10/01/2004 Expiration Date 10/31/05

Comments: _____

BREATH TEST OPERATOR CERTIFICATION RECORD

Certificate Number 13949
Operator Name DAVIDSON, JAMES G
Agency DPS PALMVIEW
Address 2401 N. MOOREFIELD R
PALMVIEW, TX, TEXAS 78572
Service Code THP
Tech Supv Area 037

PRACTICAL EXAMINATIONS

Instrument INTOXILYZER 5000


Practical Test 1

1. PASS
2. PASS
3. PASS
4. PASS
5. PASS

Practical Test 2

1. NOT TAKEN
2. NOT TAKEN
3. NOT TAKEN
4. NOT TAKEN
5. NOT TAKEN

Practical Exam PASS**WRITTEN EXAMINATION****Written Exam PASS**

Effective Date 09/23/2003
Certification Status CERTIFICATION RENEWED
Technical Supervisor 
Mariamela Martinez

OFFICE OF THE SCIENTIFIC DIRECTOR

Renewal Date 09/23/2003 Expiration Date 10/31/04

Comments: _____

BREATH TEST OPERATOR CERTIFICATION RECORD

Certificate Number 13949
Operator Name DAVIDSON, JAMES G
Agency DPS MCALLEN
Address 1414 N. BICENTENNIAL
MCALLEN, TEXAS 78501-
Service Code THP
Tech Supv Area 037

PRACTICAL EXAMINATIONS

Instrument INTOXILYZER 5000

Practical Test 1

1. PASS
2. PASS
3. PASS
4. PASS
5. PASS


Practical Test 2

1. NOT TAKEN
2. NOT TAKEN
3. NOT TAKEN
4. NOT TAKEN
5. NOT TAKEN

Practical Exam PASS

WRITTEN EXAMINATION

Written Exam N/R

Effective Date 10/01/2002
Certification Status CERTIFICATION RENEWED
Technical Supervisor 

Mariana Martinez

OFFICE OF THE SCIENTIFIC DIRECTOR

Renewal Date 10/01/2002 Expiration Date 10/31/03

Comments: _____

BREATH TEST OPERATOR CERTIFICATION RECORD

Certificate Number 13949
Operator Name DAVIDSON, JAMES G
Agency DPS MCALLEN
Address 1414 N. BICENTENNIAL
MCALLEN, TEXAS 78501-
Service Code THP
Tech Supv Area 037

PRACTICAL EXAMINATIONS

Instrument INTOXILYZER 5000


Practical Test 1

1. PASS
2. PASS
3. PASS
4. PASS
5. PASS

Practical Test 2

1. NOT TAKEN
2. NOT TAKEN
3. NOT TAKEN
4. NOT TAKEN
5. NOT TAKEN

Practical Exam PASS**WRITTEN EXAMINATION****Written Exam PASS**

Effective Date 09/27/2001
Certification Status CERTIFICATION RENEWED
Technical Supervisor 
Marianela Martinez

OFFICE OF THE SCIENTIFIC DIRECTOR

Renewal Date 09/27/2001 Expiration Date 10/31/02

Comments: _____

Texas Department of Public Safety

BREATH TEST OPERATOR CERTIFICATION RECORD

Certificate Number 13949
Operator Name DAVIDSON, JAMES G
Agency DPS MCALLEN
Address 1414 N. BICENTENNIAL
MCALLEN, TEXAS 78501-
Service Code THP
Tech Supv Area 037

PRACTICAL EXAMINATIONS

Instrument INTOXILYZER 5000

Practical Test 1

1. PASS
2. PASS
3. PASS
4. PASS
5. PASS


Practical Test 2

1. NOT TAKEN
2. NOT TAKEN
3. NOT TAKEN
4. NOT TAKEN
5. NOT TAKEN

Practical Exam PASS

WRITTEN EXAMINATION

Written Exam N/R

Effective Date 09/30/2000
Certification Status CERTIFICATION RENEWED
Technical Supervisor 
Mariana Martinez

OFFICE OF THE SCIENTIFIC DIRECTOR

Renewal Date 09/30/2000 Expiration Date 10/31/01

Comments: _____

Texas Department of Public Safety

BREATH TEST OPERATOR CERTIFICATION RECORD

Certificate Number 13949
Operator Name DAVIDSON, JAMES G
Agency DPS MCALLEN
Address 1414 N. BICENTENNIAL
MCALLEN, TEXAS 78501-
Service Code THP
Tech Supv Area 037

PRACTICAL EXAMINATIONS

Instrument INTOXILYZER 5000

Practical Test 1

1. PASS
2. PASS
3. PASS
4. PASS
5. PASS

Practical Test 2

1. NOT TAKEN
2. NOT TAKEN
3. NOT TAKEN
4. NOT TAKEN
5. NOT TAKEN

Practical Exam PASS

WRITTEN EXAMINATION

Written Exam PASS

Effective Date 09/30/1999
Certification Status CERTIFICATION RENEWED
Technical Supervisor Mary Ann Ramirez
Mary Ann Ramirez

OFFICE OF THE SCIENTIFIC DIRECTOR

Renewal Date 09/30/1999 Expiration Date 10/31/00

Comments: _____

Texas Department of Public Safety

BREATH TEST OPERATOR CERTIFICATION RECORD

Certificate Number 13949
Operator Name DAVIDSON, JAMES G
Agency DPS MCALLEN
Address 1414 N. BICENTENNIAL
MCALLEN, TEXAS 78501-
Service Code THP
Tech Supv Area 037

PRACTICAL EXAMINATIONS
Instrument INTOXILYZER 5000

Practical Test 1	Practical Test 2
1. PASS	1. NOT TAKEN
2. PASS	2. NOT TAKEN
3. PASS	3. NOT TAKEN
4. PASS	4. NOT TAKEN
5. PASS	5. NOT TAKEN

Practical Exam PASS

WRITTEN EXAMINATION

Written Exam N/R

Effective Date 10/27/1998
Certification Status CERTIFICATION RENEWED
Technical Supervisor Mary Ann Ramirez
MaryAnn Ramirez

OFFICE OF THE SCIENTIFIC DIRECTOR

Renewal Date 10/27/1998 Expiration Date 10/31/99

Comments: _____

Texas Department of Public Safety

BREATH TEST OPERATOR CERTIFICATION RECORD

Certificate Number 13949
Operator Name DAVIDSON, JAMES G
Agency DPS MCALLEN
Address 1414 N. BICENTENNIAL
MCALLEN, TEXAS 78501-
Service Code THP
Tech Supv Area 037

PRACTICAL EXAMINATIONS

Instrument INTOXILYZER 5000

Practical Test 1

1. PASS
2. PASS
3. PASS
4. PASS
5. PASS

Practical Test 2

1. NOT TAKEN
2. NOT TAKEN
3. NOT TAKEN
4. NOT TAKEN
5. NOT TAKEN

Practical Exam PASS

WRITTEN EXAMINATION

Written Exam PASS

Effective Date 09/30/97
Certification Status CERTIFICATION RENEWED
Technical Supervisor Kathy Q. Kokernot
Kathy Q Kokernot

OFFICE OF THE SCIENTIFIC DIRECTOR

Renewal Date 09/30/97 Expiration Date 10/31/98

Comments: _____

Texas Department of Public Safety

BREATH TEST OPERATOR CERTIFICATION EXAMINATION RECORD

139-4 Davidson, James G G

CERT # OPERATOR TECH. SUP. CODE

Texas Dept of Public Safety 1630 N. 77 Sunshine Strip Harlingen, TX 78550

DEPARTMENT OR AGENCY MAILING ADDRESS

THP Harlingen

DUTY STATION IF DIFFERENT

UNKNOWN SOLUTIONS LAB EXAMINATIONS

DATE <u>04-10-97</u>	INSTRUMENT <u>Innovative 5000</u>
Predicted	Test Results
<u>0. 100</u>	<u>0. 100</u>
<u>0. 110</u>	<u>0. 113</u>
<u>0. 110</u>	<u>0. 112</u>
<u>0. 110</u>	<u>0. 113</u>
<u>0. 110</u>	<u>0. 114</u>

☒ PASSED ☐ FAILED UNKNOWN Test 1 Test 2

☐ FAILED ON PAPER WORK ☐ ☐

WRITTEN EXAMINATION

DATE 4-11-97 ☒ PASSED ☐ FAILED

☒ GIVEN REASON _____

☐ NOT GIVEN

☐ NOT REQUIRED

☒ CERTIFIED ☐ RECERTIFIED ☐ INELIGIBLE

[Signature]

TECHNICAL SUPERVISOR

☒ CERTIFIED 4/11/97 to 10/31/97

DATE EXPIRATION DATE

☐ RECERTIFIED _____ to _____ ☒ MASTER RECORD SUBMITTED

DATE EXPIRATION DATE

☐ INELIGIBLE _____ to _____ ☐ MASTER RECORD UPDATED

DATE EXPIRATION DATE

Richard Baxter

SCIENTIFIC DIRECTOR OR DESIGNATE

BREATH TEST OPERATOR LABORATORY WORKSHEET

STUDENT'S NAME Last First MI

SCHOOL LOCATION

SCHOOL DATES MO DAY - MO DAY YR

REFERENCE ANALYSIS RESULTS

	PRED.	RESULT
1	0.100	0.097
2	0.100	0.098
3	0.100	0.097
4	0.100	0.097
5	0.100	0.098
6	0.100	0.098
7	0.100	0.097
8	0.100	0.098
9	0.100	0.096
10	0.100	0.097

	PRED.	RESULT
11	0.100	0.096
12	0.100	0.100
13	0.100	0.100
14	0.100	0.100
15	0.100	0.100
16	0.100	0.101
17	0.100	0.102
18	0.100	0.102
19	0.100	0.103
20	0.100	0.103

	PRED.	RESULT
21	0.100	0.102
22	0.100	0.102
23	0.100	0.102
24	0.100	0.103
25	0.100	0.103
26	0.100	0.102
27	0.100	0.100
28	0.100	0.100
29	0.100	0.101
30	0.100	0.101

CONTROLLED DRINKING EXERCISE

ALCOHOL CONCENTRATION

TIME OF TEST	SUBJECT'S INITIALS	BREATH #1	BREATH #2	DIFF.
1504	RSS	.078	.085	.007
1511	JSS	.114	.111	.003
1518	SSM	.103	.113	.010
1525	M.W.	.094	.095	.001
1532	JDO	.061	.055	.006

ALCOHOL CONCENTRATION

TIME OF TEST	SUBJECT'S INITIALS	BREATH #1	BREATH #2	DIFF.

LABORATORY EXAMINATION

TEST 1

SAMPLE	REFERENCE ANALYSIS RESULT	PREDICTED
#1	0.100	0.100
#2	0.113	0.110
#3	0.112	0.110
#4	0.113	0.110
#5	0.114	0.110

TEST 2

SAMPLE	REFERENCE ANALYSIS RESULT	PREDICTED