

**WB- PT / INR**

**Hemochron Signature Elite Daily Maintenance and Quality Control Log**

YEAR: \_\_\_\_\_

INSTRUMENT ID # : \_\_\_\_\_

MONTH: \_\_\_\_\_

SITE: \_\_\_\_\_

DAY	TIME	CLEAN EXTERIOR 10% BLEACH	TEMP RESULT 37± 1° C	ELECTRONIC QC		Liquid Normal QC	Liquid Abnormal QC	CARTRIDGE LOT # : EXP. DATE :	OPERATOR ID	COMMENTS & CORRECTIVE ACTIONS	LAB REVIEW ID / DATE
				LEVEL Normal 30±1 SEC	LEVEL Abnormal 300±1 SEC	Lot # : Exp. Date: Range:	Lot # : Exp. Date: Range:				
1											
2											
3											
4											
5											
6											
7											
8											
9											
10											
11											
12											
13											
14											
15											
16											
17											
18											
19											
20											
21											
22											
23											
24											
25											
26											
27											
28											
29											
30											
31											

KEY: NIU = NOT IN USE      NOTE: CLEAN / TEMP PROBE / ELECTRONIC QC / LIQUID QC PERFORMED EVERY 8 HOURS.