



# Temperature Log for Refrigerator - Celsius

## Days 1-15

Month/Year: \_\_\_\_\_ VFC Pin: \_\_\_\_\_

Facility Name: \_\_\_\_\_

### Monitor temperatures closely!

1. Write your initials below in "Staff Initials" and note the time in "Exact Time."
2. Record the maximum and minimum temperatures at the beginning of each clinic day. Max and min are the highest and lowest temps reached since the last time you checked them. Documenting the current temperature is optional.
3. If any out-of-range temps, see instructions to the right.
4. Keep this documentation, data logger logs, and vaccine storage troubleshooting records for three years.

### Take action if temp is out of range—too warm (above 8°C) or too cold (below 2°C).

1. Label the vaccines "Do Not Use". Store vaccine in proper conditions as quickly as possible. Do not discard vaccines unless directed to by the Oregon Immunization Program.
2. Call the Oregon VFC Helpdesk at 971-673-4832 to determine vaccine viability and next steps.
3. Document the event on your clinic's Vaccine Storage Troubleshooting Record.

Day of Month	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15
Staff Initials															
Exact Time															
Maximum Temp															
Minimum Temp															
<b>Danger! Temperatures above 8°C are too warm! Call the VFC Helpdesk (971-673-4832) immediately!</b>															
<b>Current Temperature</b>	8°C														
	7°C														
	6°C														
	5°C														
	4°C														
	3°C														
	2°C														
<b>Danger! Temperatures below 2°C are too cold! Call the the VFC Helpdesk (971-673-4832) immediately!</b>															



Optional



# Temperature Log for Refrigerator - Celsius

## Days 16-31

Month/Year: \_\_\_\_\_ VFC Pin: \_\_\_\_\_

Facility Name: \_\_\_\_\_

### Monitor temperatures closely!

1. Write your initials below in "Staff Initials" and note the time in "Exact Time."
2. Record the maximum and minimum temperatures at the beginning of each clinic day. Max and min are the highest and lowest temps reached since the last time you checked them. Documenting the current temperature is optional.
3. If any out-of-range temps, see instructions to the right.
4. Keep this documentation, data logger logs, and vaccine storage troubleshooting records for three years.

### Take action if temp is out of range—too warm (above 8°C) or too cold (below 2°C).

1. Label the vaccines "Do Not Use". Store vaccine in proper conditions as quickly as possible. Do not discard vaccines unless directed to by the Oregon Immunization Program.
2. Call the Oregon VFC Helpdesk at 971-673-4832 to determine vaccine viability and next steps.
3. Document the event on your clinic's Vaccine Storage Troubleshooting Record.

Day of Month	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31
Staff Initials																
Exact Time																
Maximum Temp																
Minimum Temp																
Danger! Temperatures above 8°C are too warm! Call the VFC Helpdesk (971-673-4832) immediately!																
<b>Current Temperature</b>	8°C															
	7°C															
	6°C															
	5°C															
	4°C															
	3°C															
	2°C															
	Danger! Temperatures below 2°C are too cold! Call the VFC Helpdesk (971-673-4832) immediately!															



Optional