



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center – WO66-G609  
Silver Spring, MD 20993-0002

March 19, 2015

Navilyst Medical Inc.  
Ms. Wanda Carpinella  
Director, Regulatory Affairs  
26 Forest Street  
Marlborough, Maine 01752

Re: K150448

Trade/Device Name: NMI PICC III and NMI PICC IV  
Regulation Number: 21 CFR 880.5970  
Regulation Name: Percutaneous, implanted, long-term intravascular catheter  
Regulatory Class: II  
Product Code: LJS  
Dated: February 18, 2015  
Received: February 20, 2015

Dear Ms. Carpinella:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

 Tina  
Kiang -S

for Erin I. Keith, M.S.

Director

Division of Anesthesiology, General Hospital,  
Respiratory, Infection Control and Dental Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K150448

Device Name

NMI PICC III and NMI PICC IV

Indications for Use (Describe)

The NMI PICC III and NMI PICC IV are indicated for short- or long-term peripheral access to the central venous system for intravenous therapy, including but not limited to, the administration of fluids, medications and nutrients, the sampling of blood, for central venous pressure monitoring and for power injection of contrast media.

### Maximum Power Injection Flow Rate

Description	Flow Rate
3F Single Lumen - 55cm Length	1 mL/sec
4F Single Lumen - 55cm Length	3.5 mL/sec
5F Single Lumen - 55cm Length	5 mL/sec
5F Dual Lumen - 55cm Length	4 mL/sec
6F Dual Lumen - 55cm Length	5 mL/sec
6F Triple Lumen - 55cm Length	6 mL/sec

The NMI PICC III with PASV Valve Technology is indicated for short- or long-term peripheral access to the central venous system for intravenous therapy, including but not limited to, the administration of fluids, medications and nutrients, the sampling of blood, for central venous pressure monitoring and for power injection of contrast media.

### Maximum Power Injection Flow Rate

Description	Flow Rate
3F Single Lumen - 55cm Length	1 mL/sec
4F Single Lumen - 55cm Length	3.5 mL/sec
5F Single Lumen - 55cm Length	5 mL/sec
5F Dual Lumen - 55cm Length	4 mL/sec
6F Dual Lumen - 55cm Length	5 mL/sec
6F Triple Lumen - 55cm Length	6 mL/sec

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRASStaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

**510(K) SUMMARY FOR THE NMI PICC III AND NMI PICC IV**

510(K) #: **K150448**

Date prepared: February 18, 2015

**A. Sponsor**

Navilyst Medical, Inc  
26 Forest Street  
Marlborough, MA 01752

**B. Contact**

Wanda Carpinella  
Director, Regulatory Affairs  
508-658-7929

Gary Barrett  
Vice President, Regulatory Affairs  
508-658-7940

**C. Device Name**

Trade Name: NMI PICC III and NMI PICC IV  
Common/Usual name: Peripherally Inserted Central Catheter (PICC)  
Classification Name: Short and Long-Term Intravascular Catheter  
21CFR§880.5970, Class II  
Classification Panel: General Hospital

**D. Predicate Device**

Common/Usual name: Peripherally Inserted Central Catheter (PICC)  
Classification Name: Short and Long-Term Intravascular Catheter  
21CFR§880.5970, Class II  
Classification Panel: General Hospital  
Premarket Notification: K140266 (NMI PICC IV), K131942 (NMI PICC III)

**E. Device Description**

**Intended Use**

The NMI PICC IV is indicated for short- or long-term peripheral access to the central venous system for intravenous therapy, including but not limited to, the administration of fluids, medications and nutrients, the sampling of blood, for central venous pressure monitoring and for power injection of contrast media.

Maximum Power Injection Flow Rate	
Description	Flow Rate
3F Single Lumen – 55cm length	1 mL/sec
4F Single Lumen – 55cm length	3.5 mL/sec
5F Single Lumen – 55cm length	5 mL/sec
5F Dual Lumen – 55 cm length	4 mL/sec
6F Dual Lumen – 55cm length	5 mL/sec
6F Triple Lumen – 55cm length	6 mL/sec

The NMI PICC III with PASV Valve Technology is indicated for short- or long-term access to the central venous system for intravenous therapy, including but not limited to, the administration of fluids, medications and nutrients, the sampling of blood, and for power injection of contrast media.

Maximum Power Injection Flow Rate	
Description	Flow Rate
3F Single Lumen – 55cm length	1 mL/sec
4F Single Lumen – 55cm length	3.5 mL/sec
5F Single Lumen – 55cm length	5 mL/sec
5F Dual Lumen – 55cm length	4 mL/sec
6F Dual Lumen – 55cm length	5 mL/sec
6F Triple Lumen – 55cm length	6 mL/sec

#### F. Summary of Similarities and Differences in Technological Characteristics and Performance

The proposed device has similar design, components and technological characteristics as the predicate intravascular catheters; the only difference between the predicate and proposed devices is the type of ink used on the markings of the catheter shaft. Biocompatibility testing per ISO 10993-1 demonstrates that the new ink on the proposed device does not affect safety and/or effectiveness.

Similarities between the proposed and predicate devices include:

- intended for short- or long-term peripheral access to the central venous system for intravenous therapy, blood sampling, for central venous pressure monitoring and power injection of contrast media.
- available in single and multi-lumen configurations in a wide range of sizes from 3F to 6 F outside catheter diameter;
- rated for maximum power injector settings up to 325 psi
- rated for maximum power injection flow rate up to 6 ml/second based on model; and
- available kitted with a range of procedural accessories for user convenience and,
- demonstrate resistance to blood components (platelet and thrombus) accumulation.

#### G. Performance Data

The NMI PICC III and NMI PICC IV are substantially equivalent to Navilyst predicate devices based on comparison of technological characteristics and the results of non-clinical tests which included the performance evaluation and biocompatibility testing conducted in accordance with the following FDA guidance documents, international standards, and testing which included:

- EN ISO 10555-1:2009, *Sterile, Single use intravascular catheters – Part 1: General Requirements*
- EN ISO 10555-3:1997 Corrigendum 1:2002, *Sterile, Single-Use Intravascular Catheters – Part 3: Central Venous Catheters*
- FDA’s “*Guidance on Premarket Notification [510(K)] Submissions for Short-Term and Long-Term Intravascular Catheters dated March 16, 1995*”
- FDA Blue Book Guidance G95-1 “*Use of International Standard ISO 10993-1, “Biological Evaluation of Medical Devices Part 1: Evaluation and Testing”*”

#### H. Conclusion

The results of the non-clinical testing and a comparison of similarities and differences demonstrate that the proposed and predicate devices are substantially equivalent.