Certificate of Analysis List For request number 493948

Catalog	Lot	Related	Related	
Number	Number	Catalog	Lot	
Entered	Entered	Number	Code	Description
1403927	2230	1403969	2256	CyaniVer® 3 Cyanide Reagent
Powder Pil	lows			
1403927	2240	1403969	2256	CyaniVer® 3 Cyanide Reagent
Powder Pil	lows			

Total Enclosures: 2



P.O.Box 389 Loveland, CO 80539 (970) 669-3050

Certificate of Analysis This is a Component of 1403969 lot 2256

Page 1

COMMODITY: CyaniVer® 3 Cyanide Reagent Powder Pillows

COMMODITY NUMBER: 14039-27

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: A2230

8/17/2012

8/21/2012

TEST	SPECIFICATIONS	RESULTS
0.05 mg/L Cyanide standard measured vs.deionized water.	0.405 to 0.459 abs	0.4190 abs
pH of 0.4 g of Cat. No. 14039 dissolved in 50 mL of deionized water.	6.65 to 7.05	6.890
Reagent blank measured vs. deionized water.	0 to 0.017 abs	0.0160 abs

The expiration date is Aug 2017

A EUT



P.O.Box 389 Loveland, CO 80539 (970) 669-3050

Certificate of Analysis This is a Component of 1403969 lot 2256

Page 1

COMMODITY: CyaniVer® 3 Cyanide Reagent Powder Pillows

COMMODITY NUMBER: 14039-27

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: A2240

8/27/2012

8/28/2012

TEST	SPECIFICATIONS	RESULTS
0.05 mg/L Cyanide standard measured vs.deionized water.	0.405 to 0.459 abs	0.4050 abs
pH of 0.4 g of Cat. No. 14039 dissolved in 50 mL of deionized water.	6.65 to 7.05	6.860
Reagent blank measured vs. deionized water.	0 to 0.017 abs	0.0170 abs

The expiration date is Aug 2017

A EUT