





97 Broadway NEDLANDS 6009 Western Australia PO Box 3258, Broadway Nedlands WA 6009 Phone: (08) 9386 3304 Facsimile: (08) 9386 3202 Internet: http://www.daa.com.au ABN 68 009 304 956

Certificate of Analysis

Reference Material: AUOH-4

	95% Confid	ence Interval	Inter-Laboratory	Inter-Unit		
Recommended Value	Lower Limit	Upper Limit	Standard Deviation	Standard Deviation		
1.3140	1.2814	1.3466	0.0564	0.0083		

Table 1 Recommended Values and 95% Confidence Intervals of Gold Concentration (in ppm, or $\mu g/g$). Based on the results of 14 laboratories.

1. Date of Certification

January 2008

2. Certificate Status

Original

3. Supplier of Reference Material

Western Mineral Standards 54 Helen Street Bellevue 6056 Western Australia

4. Prepared and Certified by

Dr John Henstridge Principal Consultant Statistician Data Analysis Australia 97 Broadway Nedlands, WA, 6009

Dr John Henstridge

WMS/8 C Reference Material: AUOH-4

5. Available Packaging

This reference material is packed in 2kg PE screw cap jars, labelled with the identification code, the expected value and standard deviation of the gold concentration, health warning and the manufacturer's contact details. They are also available in cartons of six jars, and each carton will include this certificate of analysis.

6. Origin of Reference Material

This reference material is an oxide gold ore from the Mt Magnet region in Western Australia.

The reference material has been well mixed. The entire batch was tested to ensure that the gold was evenly distributed throughout the material.

7. Intended Use

Gold Assay standards provided with associated geochemistry data.

8. Instructions for Use

Weigh out the usual quantity used in the analysis for total gold by the normal procedure. Homogeneity testing has shown that consistent results can be obtained for gold when using samples of 30 to 40 grams.

9. Method of Test Material Preparation

Two tonnes of the reference material was discharged from a homogenous mixer through a vibrasonic sieve into large bags. A one-kilogram sample was taken after every 200 kilograms of the material was discharged. These one-kilogram samples were numbered in the order of being discharged, then divided into four subsamples of approximately 250 grams.

These subsamples were placed into zip lock bags, and labelled by Data Analysis Australia personnel with a random identifier, prior to being sent to sixteen laboratories for testing.

The laboratories participating in this certification process were sent either two or three of the 250 gram subsamples. The laboratories were also sent a number of 250 gram subsamples from three other reference materials that were being analysed for certification. An indication of the expected gold concentration was given in broad terms, with a range of three parts per million.

Reference Material: AUOH-4

10. Homogeneity Tests

The homogeneity test was incorporated into the methodology by ensuring that each laboratory received at least two subsamples of the reference material, taken from samples spaced out across the ordering (see Analytical Methodology). There was no evidence that this reference material was non-homogenous across the one-kilogram samples.

11. Analytical Methodology

This reference material was analysed for certification along with three other reference materials, which allows the use of more powerful statistical techniques in the certification. Performing the certification in this manner also provides the following advantages:

- "Blinding" of the laboratories, in that they did not know which reference material each sample corresponded to, so the laboratory did not have prior knowledge as to what result they should obtain. The inclusion of the expected range of the gold concentration was to assist the laboratories in their analysis, and did not provide enough information for the laboratories to know which reference material was in each sample.
- Improved ability to detect bias between laboratories, such as if one laboratory gives results that are consistently higher/lower than other laboratories for all standards; and
- Ability to detect bias in the results between laboratories for each standard separately.

Sixteen laboratories were randomly allocated an identifying letter from A to P. The set of subsamples was designed in such a way that each laboratory received at least two subsamples of each reference material. These subsamples were taken from different ordered samples in such a way to eliminate the confounding of laboratory against the order of the sample. Table 2 shows the set of subsamples received by each laboratory, where AUOH-4 has been labelled as reference material A.

The subsamples were labelled with random identifiers so the laboratories were unaware of the reference material and ordered one kilogram sample that the subsample came from. This provided a means of determining whether a laboratory had an overall bias or a problem of precision.

Reference Material: AUOH-4

		Ordered One Kilogram Sample								
Lab	1	2	3	4	5	6	7	8	9	10
A	D	Α	С	С	Α	С	В	В	D	Α
В	С		В	В		ВD	Α	A D	С	D
C	ВD		Α	Α	D	A C		С	В	С
D	A C	D	D		С	В	D	В	Α	В
E	В	С	С	D	В	Α	С	Α	D	Α
F	A D	В	В	CD	Α		ВD		С	
\mathbf{G}	С	Α	A D	ВС			A C		В	D
H	В		С	ΑВ	D		В	D	A D	С
I	Α	D	В	Α	С	D	Α	С	С	В
J		С	Α		В	CD	D	В	ВD	Α
K		В	D		A D	ВС	С	Α	A C	
L		A D	С	D	С	АВ	В	D	В	
M	D	С	В	С	В	Α	Α	С	Α	D
N	С	BD	Α	В	A D			В		CD
O	В	A C		Α	С		D	A D		ВС
P	Α	В	D	D	В	D	С	С		АВ

Table 2 Set of subsamples by laboratory. The reference materials being tested have been labelled A through to H, with AUOH-4 labelled A.

12. Calculation of Certified Values

Results for Gold via Fire Assay were returned by 16 laboratories. These results were used to calculate the recommended values and 95% confidence intervals for gold concentrations in each of the materials.

The Analysis of Variance (ANOVA) technique was used to assess the homogeneity of the reference materials across the one-kilogram samples and also to identify outliers in the gold concentration. Outliers were detected and removed by comparing individual results, sample results and overall laboratory results. Once the outliers were removed and the homogeneity of the reference material was confirmed, the remaining results were used to calculate the certified values.

Reference Material: AUOH-4

DATA ANALYSIS AUSTRALIA PTY

The results from Laboratories E and G showed a high level of variability across all reference materials. All results from these laboratories were therefore excluded from the analysis. For this reference material, the remaining fourteen laboratories returned a total of 68 results across 34 subsamples. Four results were removed as outliers. A summary of the results from the laboratories is given in Table 3, showing average values with and without outliers. A laboratory result marked with an asterisk signifies that some of that laboratory's results were identified as outliers and removed prior to the calculation of the recommended value, confidence limits and standard deviation.

The recommended value was calculated as an average of the laboratory average gold concentrations, and the confidence interval was created using the average and standard deviation of the laboratory averages. The confidence interval is calculated using the following formula:

$$X \pm (t * s / \sqrt{n})$$

where X is the average and s is the standard deviation of the laboratory averages, n is the number of laboratory averages used and t is the 0.025 critical value of the Student's t-distribution. The recommended values and confidence limits are given in Table 1 on the first page of this certificate.

	Average Gold Concentration (µg/g)				
Laboratory	Fire Assay (Outliers Removed)	Fire Assay			
A	1.2833	1.2833			
В	1.3525	1.3525			
C	1.2817	1.2817			
D	1.3500	1.3500			
E*	-	0.9711			
F	1.3250	1.3250			
G*	-	0.7117			
Н	1.3125	1.3125			
I	1.3250	1.3250			
J*	1.2200	1.1300			
K	1.2967	1.2967			
L	1.2550	1.2550			
M	1.4583	1.4583			
N	1.3500	1.3500			
O	1.2733	1.2733			
P	1.3125	1.3125			

Table 3 Average gold concentrations (µg/g) by laboratory and method. A * indicates that outliers were removed.

Reference Material: AUOH-4

DATA ANALYSIS AUSTRALIA LTD

The scale of inhomogeniety was determined through the Analysis of Variance and is presented in Table 1 as the Inter-Unit Standard Deviation. In determining the likely precision of a single assay using this standard, it is necessary to combine this with the measurement error that may depend upon laboratory practice and should be determined for each laboratory through the analysis of duplicates. The combining of standard deviations should be done through the summing of the corresponding variances.

Note that when making comparisons between laboratories, between batches or between duplicates the appropriate consideration of errors should be used. When in doubt appropriate statistical guidance should be obtained.

13. Notice

The reference material AUOH-4 and this certificate have been prepared with the appropriate care and attention. This certificate conforms to the requirements outlined in the ISO Guide 35, Certification of reference materials – Statistical and General Principles. The Purchaser by receipt hereof releases and indemnifies Data Analysis Australia and Western Mineral Standards from and against all liability and costs arising from the use of this material.

Reference Material: AUOH-4

14. Participating Laboratories

The laboratories that agreed to participate are shown in Table 4.

Laboratory	State/Province/City	Country	
Acme Analytical Laboratorie	·S		
Ltd	British Columbia	Canada	
Activation Laboratories	Ontario	Canada	
ALS Chemex Canada	British Columbia	Canada	
ALS Chemex Perth	Western Australia	Australia	
Amdel Adelaide	South Australia	Australia	
Amdel NZ	Macraes	New Zealand	
Amdel Perth	Western Australia	Australia	
Ammtec	Western Australia	Australia	
Anglo Research	Johannesburg	South Africa	
Genalysis (Perth)	Western Australia	Australia	
ITS Jakarta	Jakarta	Indonesia	
SGS Kalgoorlie	Western Australia	Australia	
SGS Townsville	Queensland	Australia	
SGS Welshpool	Western Australia	Australia	
SGS West Wyalong	New South Wales	Australia	
Ultra Trace	Western Australia	Australia	

Table 4 Participating laboratories. (Not related to the order of laboratories in other tables.)

Reference Material: AUOH-4